

CLINICAL EVALUATION SUMMARY - CES STR F 01

This summary has been compiled from the results of a number of returned Clinical Evaluation forms completed by both prosthetists and patients, and shown in an abbreviated form overleaf. It is an attempt to give an overview of the product based on our experience to date and needs to be read in conjunction with the product literature supplied by the manufacturer

STREIFENEDER – DYNAMIC SACH

WARRANTY PERIOD – 1 YEAR
WEIGHT LIMIT – 125kg



EVALUATION SUMMARY

This foot has proved to be a good alternative prescription for a patient requiring a Dynamic SACH Foot similar to that most commonly prescribed in the UK. It appears to have a slightly stiffer forefoot, which some patients may benefit from, along with the reported improvement in energy return.

INDICATIONS	CONTRAINDICATION
Exoskeletal and water activity prostheses.	Patients weighing more than 125kg.
A desire for an improved cosmetic appearance over alternative foot prescriptions.	Patients who require a high degree of compliance with uneven terrain.
All patients who would benefit from a SACH Dynamic foot design.	Hi activity/impact levels of use.

EVALUATION PATIENTS

PATIENT DETAILS

Patient 1	Transtibial	105kg	41 year old male	Prosthetist	Sigam F
Patient 2	Transtibial	67kg	57 year old female	Administration Assistant	Sigam F
Patient 3	Transtibial	90kg	41 year old male	Unemployed	Sigam F
Patient 4	Transtibial	90kg	69 year old male	Retired	Sigam C

